


This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo, copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

UNEDITED PROOF

# Accepted Manuscript

**Title:** Determination of quercetin in dietary supplements by isocratic RP-HPLC method

Authors: Dragica Doneva<sup>1\*</sup>, Biljana Bauer<sup>2</sup>



<sup>1</sup>*Institute of Public Health of the Republic of North Macedonia, 50 Division 6, 1000 Skopje, Republic of North Macedonia*

<sup>2</sup>*Institute of Pharmacognosy, Faculty of Pharmacy, Ss. Cyril and Methodius University, Mother Theresa 47, 1000 Skopje, Republic of North Macedonia*

DOI:

Received date: November 2021

Accepted date: December 2021

UDC:

Type of paper: Original scientific paper

Mac. Pharm. Bull. Vol. 67(2) 2021

Please cite this article as:

**Determination of quercetin in dietary supplements by isocratic RP-HPLC method**Dragica Doneva<sup>1\*</sup>, Biljana Bauer<sup>2</sup>

<sup>1</sup>*Institute of Public Health of the Republic of North Macedonia, 50 Division 6,  
1000 Skopje, Republic of North Macedonia*

<sup>2</sup>*Institute of Pharmacognosy, Faculty of Pharmacy, Ss. Cyril and Methodius University,  
Mother Theresa 47, 1000 Skopje, Republic of North Macedonia*

**Abstract**

Quercetin is a natural flavonoid found abundantly in vegetables and fruits. It shows antioxidant, anti-inflammatory, antihepatotoxic, antiallergic, antidiabetic and antiviral activity. Increasingly popular dietary supplements containing quercetin require critical examination of their quality. The aim of this study was to develop a simple and accurate HPLC method for quercetin determination in dietary supplements. Chromatographic separation was achieved by isocratic method, using a Purospher STAR<sup>®</sup> RP-18 reverse-phase column (150 x 4.6 mm i.d., particle size 5 µm), a mobile phase composed of acetonitrile and water (acidified to pH 3.0), in a ratio of 30:70 (V/V), run at a flow rate of 1.1 mL/min. The column temperature was kept at 30 °C. The DAD detector was set at 257 nm and 375 nm. The injection volume was 20 µL. Isopropanol was used as a solvent. The method was validated by determining system suitability, specificity, linearity, range, limit of detection and quantification, accuracy, precision, and robustness. It is characterized by simple preparation, good precision (RSD < 2%) and good analytical yield (100.09% and 100.29% at 257 nm and 375 nm, respectively). The limit of detection and quantification were 0.00071 µg/mL and 0.00215 µg/mL at 257 nm, and 0.00078 µg/mL and 0.00236 µg/mL at 375 nm, respectively. The system suitability test showed that method performance is similar at both wavelengths. This method can be recommended for routine analysis of dietary supplements containing quercetin in food quality control laboratories.

**Keywords:** quercetin, HPLC, dietary supplements, flavonoid

## Introduction

Quercetin is a natural flavonoid which belongs to the category called flavonols. It is ubiquitously present in various vegetables and fruits, like berries, lovage, capers, cilantro, dill, apples, and onions (Anand et al., 2016). It is one of the most common flavonoids found in herbal products and its biological activity is important in understanding the health benefits of these compounds (Ay et al., 2021; Lakhanpal and Rai, 2007).

The IUPAC chemical name of quercetin is 2-(3,4-dihydroxyphenyl)-3,5,7-trihydroxychromen-4-one, with a chemical formula  $C_{15}H_{10}O_7$  (Fig. 1). It is yellow in color and completely soluble in lipids and alcohol, insoluble in cold water, while sparingly soluble in hot water (Batiha et al., 2020).

Fig. 1

Quercetin has shown many pharmacological activities, such as being anticancer, antiviral, and treating allergic, metabolic, and inflammatory disorders, eye and cardiovascular diseases, and arthritis (Dabeek and Marra, 2019; Harborn and Williams, 2000; Manach et al., 2005), and as an adjunct therapy for diseases related to thrombosis (Kovačević, 2000).

Quercetin is one of the most potent antioxidants among polyphenols (Prior, 2003; Russo et al., 2000), an anticancer and neuroprotective (Chen et al., 2017; Khan et al., 2019) and is characterized by very high safety, both in animals and in humans (Dajas, 2012). Through its flavonoid ingredients and together with ascorbic acid, introduced into both the human and animal organism, enters a series of oxidative processes (Kulevanova, 2004). The beneficial effect of quercetin is believed to be due to its ability to act as a free radical scavenger and its potency to inhibit the proliferation of transformed epithelial cells. It has effect on increasing the apoptosis of mutated cells, inhibiting cancer cell carcinoma, inhibiting DNA synthesis, reducing, and modifying cell signal transduction pathways

(Erkoc et al., 2003) and is able to prevent cancer caused by oxidative stress (Bagel et al., 2012).

Quercetin is effective against various viruses. It is potent against bursting in MT2 cells caused by human T-lymphotropic virus 1 (Coelho-Dos-Reis et al., 2011). Quercetin-3-O- $\beta$ -D-glucuronide is effective against influenza A virus (Fan et al., 2011) and quercetin 7-ramnoside is effective against swine diarrhea epidemic virus (Song, 2011). It acts as a bacteriostatic agent by inhibiting bacterial growth (Wu et al., 2008) with excellent results in *Escherichia coli*, *Salmonella enteric* and *Listeria monocytogenes* (Božič et al., 2012) and helps reduce microbial growth (Andres et al., 2013).

Administration of the increasingly popular dietary supplements containing quercetin requires critical examination of quality, safety, and efficacy of these products. Although many high-quality dietary supplements exist on the market, quality concerns are in general justified. There are several studies that report that active ingredients dramatically differ from label claims and among products (Abdel-Tawab, 2018). Commonly identified problems include failure to verify that a finished batch meets product specifications for identity, purity, strength, and composition (Kapoor and Sharfstein, 2016). Quality control is one of the very important and essential steps in the manufacturing of dietary supplements as quality of products affect their safety and efficacy (Kagawad et al., 2021). Hence, the determination of the active ingredients in the finished products is required to check the compliance with the product specifications and to prove they contain the ingredients listed on the label in the declared potency and amount.

Quercetin in plants, biological materials, and pharmaceutical preparations is determined by spectrophotometric, chromatographic, and electro-migration methods (Kurzawa, 2010). HPLC technique is used for determination of quercetin in different matrices, e.g., in the red wines (Fang et al., 2007; Sakkiadi et al., 2001; Stecher et al., 2001); phyto-pharmaceutical preparations (Mesbah et al., 2005); and in human urine after oral administration of a tablet of Ginkgo biloba extract (Wang et al., 2003). To date, there is no official monograph on quercetin in the European Pharmacopoeia (Ph. Eur., 10.0, 2019).

Therefore, the aim of this study was to develop simple, accurate, and reproducible high-performance liquid chromatography (HPLC) method for routine determination of

quercetin in commercially available dietary supplements.

## Materials and methods

### *Materials and reagents*

We used commercially available capsules with a standardized content of 250 mg quercetin and 250 mg ascorbic acid, manufactured by PharmaVital GmbH, Germany, as samples. We chose this product because of its availability on the local market. In addition, the presence of vitamin C, beside quercetin, will allow us to confirm that quercetin can be accurately determined in the presence of vitamin C. Quercetin standard substance (a purity of the standard substance was 0.98) was purchased from Sigma-Aldrich (St. Louis, MO, USA). Ascorbic acid standard substance was purchased from Sigma-Aldrich (St. Louis, MO, USA). HPLC-grade acetonitrile and isopropanol were purchased from Merck (Darmstadt, Germany). Double-distilled water was used for preparation of mobile phase solutions. Phosphoric acid was supplied from Sigma-Aldrich (St. Louis, MO, USA). The volumetric glassware used in the study were class A. Prior to injection into the HPLC system, all solutions were filtered through a membrane nylon filter (0.45  $\mu\text{m}$  pore size).

### *Instrumentation*

In this study HPLC analyses were performed using an Agilent Technologies chromatographic system (Hewlett Packard, Avondale, USA) consisting of a Model 1200 Series SL pump with a Model Agilent series G-1315C DAD detector and a Model Agilent 1200 series G-1329B ALS auto sampler. Data analyses were done using Agilent Technologies HPLC 1100 software. The UV spectra of quercetin dissolved in isopropanol were obtained in the wavelength region from 200 to 400 nm using Model Lambda 12 (Perkin Elmer) UV-visible spectrophotometer. The wavelength of 257 nm and 375 nm was chosen for the quantification of quercetin.

### *Chromatographic Conditions*

Chromatographic separation was achieved using a LiChroCART<sup>®</sup> Purospher<sup>®</sup> STAR RP-18 end-capped column (150 x 4.6 mm i.d., particle size 5  $\mu\text{m}$ ). The mobile phase

was composed of mixture of acetonitrile and water acidified with o-phosphoric acid (pH 3.0) in ratio 30:70 (V/V), filtered through 0.45  $\mu\text{m}$  nylon filter. The column temperature was maintained at 30  $^{\circ}\text{C}$ , while the eluent was run at 1.1 mL/min. The injection volume was 20  $\mu\text{L}$ . Isopropanol was used as a solvent in preparation of standard and sample solution. The operating wavelength was selected by scanning a mix standard solution (quercetin and ascorbic acid in isopropanol) in the wavelength region from 200 to 400 nm using a Model Lambda 12 (Perkin Elmer) UV-visible spectrophotometer (Fig. 2).

#### *Calibration curve*

A quercetin stock solution was prepared by dissolving quercetin standard substance (25.12 mg) with isopropanol in 25 mL amber volumetric flask ( $c = 0.9851$  mg/mL). The mixture was sonicated in an ultrasonic bath for 10 min and stored at room temperature (25  $\pm 2$   $^{\circ}\text{C}$ ) during the study.

Working standard solutions were prepared daily, by dilution of quercetin stock solution with the same solvent (isopropanol) to obtain final concentrations ranging from 4.93 – 118.21  $\mu\text{g/mL}$  (4.93; 9.85; 19.70; 39.40; 98.50 and 118.21  $\mu\text{g/mL}$ ). Isopropanol was used as a blank.

#### *Sample preparation*

The following procedure was used to determine the quercetin content in capsules. Twenty capsules were accurately weighed and used to calculate the average mass of one capsule. Then, a capsule content containing 250 mg quercetin (approximately an average capsule content) was accurately weighed, transferred to an amber 100 mL volumetric flask, about 70 mL of isopropanol were added in the flask and the solution was sonicated in an ultrasonic bath for 10 min. The volume was made up to the mark with the same solvent (target concentration of 0.02 mg/mL quercetin). All solvents and solutions were filtered through a nylon membrane filter with a pore size of 0.45  $\mu\text{m}$ . The amount of quercetin per capsule was calculated using the standard calibration curves at two wavelengths (257 nm and 375 nm).

## Results and discussion

Various mobile phases and chromatographic columns have been used to optimize the chromatographic conditions of a HPLC method for determination of quercetin in dietary supplements. The selection of the mobile phase and HPLC column was done on a basis of the system suitability parameters (capacity factor, tailing factor, theoretical plate etc.). Thus, the mobile phase consisting of acetonitrile and water acidified with o-phosphoric acid (pH 3.0), 30:70 (V/V), was found optimal for isocratic determination of quercetin in capsules (Table 1).

Table 1

The wavelengths were selected by scanning the mix standard solution (quercetin and vitamin C in isopropanol) over 200 – 400 nm and the wavelengths of 257 nm and 375 nm were selected for determination of quercetin in dietary supplements also containing ascorbic acid. While quercetin exhibits maximum absorbance at 375 nm, ascorbic acid has showed very poor absorbance at this wavelength (Fig. 2). Therefore, we decided to include the additional wavelength of 257 nm (besides the wavelength of 375 nm) in order to monitor whether the proper chromatographic separation of the quercetin and ascorbic acid is achieved.

Fig. 2

### *Method validation*

We used the ICH guideline on the validation of analytical procedures to demonstrate that the developed method is suitable for its intended purpose (ICH, 2005). As the method will be used for determination of quercetin in capsules, the mandatory parameters for method validation are specificity/ selectivity, linearity, limit of detection (LOD), limit of quantification (LOQ), accuracy, precision and robustness.

*Specificity and selectivity*

Specificity is applied to provide “an exact result which allows an accurate statement on the content or potency of the analyte in a sample in the presence of components that can be expected to be present (impurities, degradants, matrix, etc.)” (ICH, 2005). The resolution between quercetin peak and vitamin C peak was 7.4 at 257 nm which proved that critical separation between these two components which elute closest to each other was achieved (at 375 nm ascorbic acid has showed poor absorption) (Fig. 3).

Fig. 3

We used typical chromatograms of standard solution, sample solution and blank solution (isopropanol) to demonstrate specificity. Specificity was demonstrated as there was no overlap of the main peak with the peaks from the blank solution, or eventually present impurities and degradants (Fig. 4 and Fig. 5).

Fig. 4

Fig. 5

Furthermore, the UV spectrum of quercetin peak, eluted at retention time of 5.98 min from the standard solution was similar with the quercetin peak from the sample solution (containing quercetin and ascorbic acid), which proved there was no interference from the ascorbic acid and other ingredients (Fig. 6).

Fig. 6

Considering the obtained results of the system suitability test, the method performance is similar at both wavelengths (257 nm and 375 nm) (Table 1). However, ascorbic acid is visible only at the wavelength of 257 nm because this component has shown poor absorbance at 375 nm.

### *Linearity and range*

To show that the results of the analytical method are directly proportional to the concentration of the quercetin, we used 6 different concentrations of standard quercetin solutions (0.0049 – 0.11821 mg/mL). By mathematical estimation of the degree of dependence, we obtain similar equations of the regression line at both wavelengths, as follows:

$$y = 64179 * x - 71.533 \text{ (257 nm)}$$

$$y = 65032 * x - 79.493 \text{ (375 nm)}$$

Together with the correlation coefficients:

$$R^2 = 1.0000 \text{ (257 nm)}$$

$$R^2 = 0.9999 \text{ (375 nm)}$$

Table 2

### *Limit of detection and limit of quantification*

The limit of detection in a single analytical procedure is defined as the amount of analyte in a sample that can be detected but does not have to be measured as an exact value. The quantitative limit of an individual analytical procedure is the amount of analyte in a sample that can be determined quantitatively with appropriate precision and accuracy. LOD and LOQ are found using the ratios of  $3.3 \sigma/S$  and  $10 \sigma/S$  respectively, where  $\sigma$  represents the standard estimation error and  $S$  is the slope (ICH, 2005).

The established LOD and LOQ for quercetin were 0.00071  $\mu\text{g/mL}$  and 0.00215  $\mu\text{g/mL}$  at 257 nm, and 0.00078  $\mu\text{g/mL}$  and 0.00236  $\mu\text{g/mL}$  at 375 nm, respectively (Table 2).

### *Accuracy*

Accuracy as a procedure for measuring uncorrected bias (system error) in analysis was performed using a standard addition method (Nethercote and Ermer, 2014). To study the accuracy of the proposed analytical method, we conducted recovery tests using the standard addition method. In order to discover whether excipients interfered with the

analysis, known amounts of standard were added to the samples and the resulting mixtures were analyzed by the proposed method. The analytical yield (recovery) was calculated at both wavelengths (Table 3).

The results obtained in the validation study (recovery within the range of 98.0 - 102.0%) show that the method accuracy is in accordance with the criteria set by Bliesner (2006).

Table 3

#### *Precision*

The precision of the analytical procedure was studied by analyzing six samples of the same sample batch at the same conditions (at the working concentration of quercetin (0.02 mg/mL) on the same day. The analysis of the samples was performed by the same analyst, using the same equipment. Intermediate precision of the analytical procedure was examined by sample analysis over three consecutive days. The precision of the analytical procedure was evaluated and expressed as a relatively standard deviation of the series of measurements (Table 4).

Table 4

The results obtained in the validation study ( $RSD \leq 2.0\%$ ) show that the method precision is in accordance with the criteria set by Bliesner (2006).

#### *Robustness*

Robustness is a measure of the ability of the analytical procedure to remain unchanged in small, but deliberate variations of the method parameters. Variations were made by changing the flow rate of the mobile phase ( $\pm 0.1$  mL/min) and the column temperature (+5 °C; -2 °C; -5 °C). The obtained values were within acceptable limits (< 2%) (Bliesner, 2006; Ph. Eur., 10.0, 2019).

From the results obtained in the study of method robustness, we can conclude that the method small and deliberate variations in flow rate and column temperature have no significant effect on system stability (Table 5).

Table 5

## Conclusion

A simple, accurate, and reproducible high-performance liquid chromatography (HPLC) method has been developed and validated for the quantification of quercetin in dietary supplements. This proposed method enables fast, direct, and simple determination of quercetin, using isopropanol as a solvent, without prior extraction from the samples. The method exhibited good linearity ( $R^2 > 0.9999$ ) over the studied concentration range (4.93 – 118.21  $\mu\text{g/mL}$ ). The proposed method is very sensitive as the established LOD and LOQ for quercetin were 0.00071  $\mu\text{g/mL}$  and 0.00215  $\mu\text{g/mL}$  at 257 nm, and 0.00078  $\mu\text{g/mL}$  and 0.00236  $\mu\text{g/mL}$  at 375 nm, respectively. It has demonstrated good intra-day and inter-day precision (relative standard deviations values were  $< 2\%$ ) and excellent method accuracy (with the recovery of 100.09% and 100.29% for quercetin, at 257 nm and 375 nm, respectively). Results from the system suitability test showed that the method performance is similar at both wavelengths (257 nm and 375 nm). The method was also successfully applied to determine quercetin in commercially available capsules containing quercetin and ascorbic acid concomitantly. It could be recommended for routine analysis of the dietary supplements containing quercetin in food quality control laboratories.

## References

Abdel-Tawab, M., 2018. Do we need plant food supplements? A critical examination of quality, safety, efficacy, and necessity for a new regulatory framework. *Planta medica* 84(06/07), 372-393. Available at: <https://doi.org/10.1055/s-0043-123764>

- Anand David, A.V., Arulmoli, R., Parasuraman, S., 2016. Overviews of biological importance of quercetin: a bioactive flavonoid. *Pharmacogn. Rev.* 10, 84-89. Available at: <https://doi.org/10.4103/0973-7847.194044>
- Andres, S., Tejido, M.L., Bodas, R., Moran, L., Prieto, N., Blanco, C. and Giraldez, F.J., 2013. Quercetin dietary supplementation of fattening lambs at 0.2% rate reduces discoloration and microbial growth in meat during refrigerated storage. *Meat Sci.* 93, 207-212. Available at: <https://doi.org/10.1016/j.meatsci.2012.08.023>
- Ay, M., Charli, A., Jin, H., Anantharam, V., Kanthasamy, A., Kanthasamy, A.G., 2017. Quercetin. In *Nutraceuticals*. Academic Press 4(1), 749-755. Available at: <https://doi.org/10.1111/jnc.14033>
- Baghel, S.S., Shrivastava, N., Baghel, R.S., Agrawal, P. and Rajput, S., 2012. A review of quercetin: antioxidant and anticancer properties. *World J. Pharm. Pharm. Sci.* 1(1), 146-160.
- Batiha, G.E.S., Beshbishy, A.M., Ikram, M., Mulla, Z.S., El-Hack, M.E.A., Taha, A.E., Algammal, A.M., Elewa, Y.H.A., 2020. The pharmacological activity, biochemical properties, and pharmacokinetics of the major natural polyphenolic flavonoid: quercetin. *Foods* 9(3), 374. Available at: <https://doi.org/10.3390/foods9030372>
- Bliesner, M.D., 2006. *Validating chromatographic methods, A practical guide*. John Wiley & Sons, Inc. Hoboken. New Jersey. ISBN: 978-0-471-74147-3
- Božič, M., Gorgieva, S. and Kokol, V., 2012. Homogeneous and heterogeneous methods for laccase-mediated functionalization of chitosan by tannic acid and quercetin. *Carbohydr. Polym.* 89, 854-864. Available at: <https://doi.org/10.1016/j.carbpol.2012.04.021>
- Chen, B.H., Park, J.H., Ahn, J.H., Cho, J.H., Kim, I.H., Lee, J.C., Won, M.H., Lee, C.H., Hwang, I.K., Kim, J.D., Kang, I.J., Cho, J.H., Shi, B.N., Kim, Y.H., Lee, Y.L., Park, S.M., 2017. Pretreated quercetin protects gerbil hippocampal CA1 pyramidal neurons from transient cerebral ischemic injury by increasing the expression of antioxidant enzymes. *Neural Regen. Res.* 12, 220-227. Available at: <https://doi.org/10.4103/1673-5374.200805>

- Coelho-Dos-Reis, J.G., Gomes, O.A., Bortolini, D.E., Martins, M.L., Almeida, M.R., Martins, C.S., Carvalho, L.D., Souza, J.G., Vilela, J.M., Andrade, M.S., Barbosa-Stancioli, E.F., 2011. Evaluation of the effects of Quercetin and Kaempferol on the surface of MT-2 cells visualized by atomic force microscopy. *J. Virol. Methods* 174, 47-52. Available at: <https://doi.org/10.1016/j.jviromet.2011.03.019>
- Dabeek, W.M., Marra, M.V., 2019. Dietary quercetin and kaempferol: Bioavailability and potential cardiovascular-related bioactivity in humans. *Nutrients* 11(10), 2288. Available at: <https://doi.org/10.3390/nu11102288>
- Dajas, F., 2012. Life or death: Neuroprotective and anticancer effects of quercetin. *J. Ethnopharmacol.* 143, 383-396. Available at: <https://doi.org/10.1016/j.jep.2012.07.005>
- Erkoc, S., Erkoc, F., Keskin, N., 2003. Theoretical investigation of quercetin and its radical isomers. *J. Mol. Struct.* 631, 141-146. Available at: [https://doi.org/10.1016/S0166-1280\(3\)00237-9](https://doi.org/10.1016/S0166-1280(3)00237-9)
- Ermer, J., Miler, J.H.M.B. 2005. *Method Validation in Pharmaceutical Analysis. A Guide to Best Practice.* WILEY-VCH Verlag GmbH & Co. KGaA. Weinheim. Available at: <https://doi.org/10.1002/3527604685>
- European Pharmacopoeia, 10<sup>th</sup> Edition, 2021. European directorate for quality of medicines, Council of Europe, Strasbourg, France.
- Fan, D., Zhou, X., Zhao, C., Chen, H., Zhao, Y., Gong, X., 2011. Anti-inflammatory, antiviral and quantitative study of quercetin-3-O-beta-D-glucuronide in *Polygonum perfoliatum* L. *Fitoterapia* 82, 805-810. Available at: <https://doi.org/10.1016/j.fitote.2011.04.007>
- Fang, F., Li, J.M., Zhang, P., Tang, K., Wang, W., Pan, Q.H., Huang, W.D., 2008. Effects of grape variety, harvest date, fermentation vessel and wine ageing on flavonoid concentration in red wines. *Food Research International* 41(1), 53-60. Available at: <https://doi.org/10.1016/j.foodres.2007.09.004>
- Harborne, J.B., Williams, C.A., 2000. Advances in flavonoid research since 1992. *Phytochemistry* 55, 481-504. Available at: [https://doi.org/10.1016/s0031-9422\(00\)00235-1](https://doi.org/10.1016/s0031-9422(00)00235-1)

- International Conference on Harmonization of Technical Requirements for Registration of Pharmaceuticals for Human Use. ICH Harmonized Tripartite Guideline: Validation of Analytical Procedures: Text and Methodology Q2 (R2), 2005.
- Kagawad, P., Gharge, S., Jivaje, K., Hiremath, S.I., Suryawanshi, S.S., 2021. Quality control and standardization of Quercetin in herbal medicines by spectroscopic and chromatographic techniques. *Future Journal of Pharmaceutical Sciences* 7(1), 1-12. <https://doi.org/10.1186/s43094-021-00327-y>
- Kapoor, A., Sharfstein, J.M., 2016. Breaking the gridlock: Regulation of dietary supplements in the United States. *Drug testing and analysis* 8(3-4), 424-430. <https://doi.org/10.1002/dta.1892>
- Khan, H., Ullah, H., Aschner, M., Cheang, W.S., Akkol, E.K., 2020. Neuroprotective Effects of Quercetin in Alzheimer's Disease. *Biomolecules* 10(1), 7-8. Available at: <https://doi.org/10.3390/biom10010059>
- Kovačević, N., 2000. Basics of pharmacognosy. Faculty of Pharmacy. University of Belgrade. Belgrade. 136.
- Kulevanova, S., 2004. Pharmacognosy: Phytochemistry and natural medicinal and aromatic raw materials. Kultura. Skopje, pp. 182.
- Kurzawa, M., 2010. Determination of quercetin and rutin in selected herbs and pharmaceutical preparations. *Analytical letters* 43(6), 993-1002. Available at: <https://doi.org/10.1080/00032710903491070>
- Lakhanpal, P., Rai, D.K., 2007. Quercetin: A versatile flavonoid. *Int. J. Med.* 2, 22-37. Available at: <https://doi.org/10.4314/ijmu.v2i2.39851>
- Manach, C., Williamson, G., Morand, C., Scalbert, A., Remesy, C., 2005. Bioavailability and bioefficacy of polyphenols in humans. I. Review of 97 bioavailability studies. *Am. J. Clin. Nutr.* 81, 230-242. Available at: <https://doi.org/10.1093/ajcn/81.1.230S>
- Mesbah, M.K., Khalifa, S.I., El-Gindy, A., Tawfik, K.A., 2005. HPLC determination of certain flavonoids and terpene lactones in selected Ginkgo biloba L. phytopharmaceuticals. *II Farmaco* 60(6-7), 583-590. Available at: <https://doi.org/10.1016/j.farmac.2005.01.014>

- Nethercote, P., Ermer, J., 2014. Analytical validation within the pharmaceutical lifecycle. *Method Validation in Pharmaceutical Analysis: A Guide to Best Practice*, 1-10. Available at: <https://doi.org/10.1002/9783527672202.ch1>
- Prior, R.L., 2003. Fruits and vegetables in the prevention of cellular oxidative damage. *Am. J. Clin. Nutr.* 78, 570-578. Available at: <https://doi.org/10.093/ajcn/78.3.570S>
- Russo, A., Acquaviva, R., Campisi, A., Sorrenti, V., Di Giacomo, C., Virgata, G., Barcellona, M.L., Vanella, A., 2000. Bioflavonoids as antiradicals, antioxidants and DNA cleavage protectors. *Cell. Biol. Toxicol.* 16, 91-98. Available at: <https://doi.org/10.023/a:1007685909018>
- Sakkiadi, A.V., Stavrakakis, M.N., Haroutounian, S.A., 2001. Direct HPLC assay of five biologically interesting phenolic antioxidants in varietal Greek red wines. *LWT-Food Science and Technology* 34(6), 410-413. Available at: <https://doi.org/10.1006/fstl.2001.0792>
- Song, J.H., Shim, J.K., Choi, H.J., 2011. Quercetin 7-rhamnoside reduces porcine epidemic diarrhea virus replication via independent pathway of viral induced reactive oxygen species. *Virology* 460, 460. Available at: <https://doi.org/10.1186/1743-422X-8-460>
- Stecher, G., Huck, C.W., Popp, M., Bonn, G.K., 2001. Determination of flavonoids and stilbenes in red wine and related biological products by HPLC and HPLC-ESI-MS-MS. *Fresenius' journal of analytical chemistry* 371(1), 73-80. Available at: <https://doi.org/10.1007/s002160100898>
- Wang, F.M., Yao, T.W., Zeng, S., 2003. Determination of quercetin and kaempferol in human urine after orally administrated tablet of ginkgo biloba extract by HPLC. *Journal of pharmaceutical and biomedical analysis* 33(2), 317-321. Available at: [https://doi.org/10.1016/s0731-7085\(03\)00255-3](https://doi.org/10.1016/s0731-7085(03)00255-3)

## Резиме

**Определување на кверцетин во додатоци на исхрана со изократски  
RP-HPLC метод**Драгица Донева\*, Билјана Бауер<sup>2</sup><sup>1</sup>Институт за јавно здравје на Република Северна Македонија, 50 Дивизија 6,  
1000 Скопје, Република Северна Македонија<sup>2</sup>Институт за фармакогнозија, Фармацевтски Факултет,  
Универзитет „Св. Кирил и Методиј“, Мајка Тереза 47,  
1000 Скопје, Република Северна Македонија**Клучни зборови:** кверцетин, HPLC, додатоци во исхрана, флавоноид

Кверцетинот е природен флавоноид кој се наоѓа во изобилство во зеленчукот и овошјето. Показува антиоксидантна, антиинфламаторна, антихепатотоксична, антиалергична, антидијабетична и антивирусна активност. Поради големата популарност на додатоците во исхраната со кверцетин, важно е да се контролира нивниот квалитет. Цел на овој труд беше да се развие едноставен и точен HPLC метод за определување на кверцетин во додатоци во исхрана. Хроматографското раздвојување беше постигнато со изократски метод, користејќи реверзно-фазна колона Purospher STAR® RP-18 (150 x 4,6 mm I.D., големина на честички 5 µm), со мобилна фаза составена од ацетонитрил и вода (закиселена со о-фосфорна киселина до pH 3,0), во сооднос од 30:70 (V/V) и проток од 1,1 mL/min. Температура на колоната беше 30 °C, а волумен на инјектирање 20 µL. DAD детекторот беше поставен на 275 nm и 375 nm. Како растворувач е користен изопропанол. Методот е валидиран со утврдување на соодветност на системот, специфичност, линеарност, опсег, лимит на детекција и квантификација, точност, прецизност и робустност. Се карактеризира со едноставна подготовка, добра прецизност (RSD < 2%) и добар аналитички принос (100,09% и 100,29% на 257 nm и 375 nm, соодветно). Границите

на детекција и квантификација се: 0,00071  $\mu\text{g/mL}$  и 0,00215  $\mu\text{g/mL}$  на 257 nm, 0,00078  $\mu\text{g/mL}$  и 0,00236  $\mu\text{g/mL}$  на 375 nm, соодветно. Тестот на соодветност на системот покажа дека методот дава споредливи резултати на двете бранови должини. Методот беше успешно применет за определување на кверцетин во диететски додаток во форма на капсули и се препорачува за рутинска анализа на кверцетин во додатоци на исхрана во лабораториите за контрола на квалитетот на храната.

Table 1. Results of system suitability test

Parameter	Results		Parameter Limit
	257 nm	375 nm	
RSD of retention time	0.09%	0.09%	< 1%
RSD of peak area	0.07%	0.05%	< 1%
Capacity factor k'	4.43	4.43	> 2
Tailing factor T	1.05	1.05	< 2
Theoretical plate	4550	4520	> 2000

Table 2. Characteristics of the linear regression analysis

Parameter	Quercetin API <sup>a</sup>	
	257 nm	375 nm
Linearity range ( $\mu\text{g/mL}$ )	4.93 – 118.21	
Slope	64179	65042
Intercept	-71.53	-79.49
Determination coefficient ( $r^2$ )	1.0000	0.9999
SE <sup>b</sup> of the intercept	13.78	15.39
SE of the slope	210.50	235.00
P-value of the slope	0.007	0.007
LOD (Limit of detection)	0.00071	0.00078
LOQ (Limit of quantification)	0.00215	0.00236

<sup>a</sup>Mean value of 3 determinations

<sup>b</sup>SE – Standard error

Table 3. Accuracy of the method

Amount added ( $\mu\text{g/mL}$ )	Amount found ( $\mu\text{g/mL}$ ) <sup>a</sup>		Recovery (%)		RSD (%)		Bias (%)	
	257 nm	375 nm	257 nm	375 nm	257 nm	375 nm	257 nm	375 nm
	16.12	16.39	16.37	101.69	101.56	0.15	0.28	1.69
18.48	18.42	18.38	99.65	99.36	0.04	0.14	-0.35	-0.64
23.21	23.09	23.39	99.49	100.67	0.05	0.05	-0.51	0.67
25.57	25.90	25.82	101.26	100.95	0.04	0.01	1.26	0.95
32.67	31.89	31.97	98.35	98.95	0.03	0.32	-1.65	-1.06
	$\bar{X}$		100.09	100.29	0.06	0.16	0.14	1.18

<sup>a</sup>  $n = 3$  determinations

Table 4. Precision of the method

Precision Parameter	% of declared content ( $\bar{X}$ )		RSD (%)		Bias (%)	
	257 nm	375 nm	257 nm	375 nm	257 nm	375 nm
Wavelength	257 nm	375 nm	257 nm	375 nm	257 nm	375 nm
Intra-assay Precision <sup>a</sup>	100.86	100.22	0.32	0.97	0.86	0.22
Intermediate Precision <sup>a</sup>						
Day 1	100.02	99.96	0.11	0.08	0.02	-0.04
Day 2	99.88	99.72	0.14	0.26	-0.12	-0.28
Day 3	99.55	99.66	0.27	0.17	-0.45	-0.34
$\bar{X}$	99.82	99.76	0.17	0.89	-0.18	-0.24

<sup>a</sup>  $n = 6$

Table 5. Results of the robustness of the method

Temperature	RSD	RSD	RSD	RSD	RSD	RSD	RSD	RSD
	(%)	(%)	(%)	(%)	(%)	(%)	(%)	(%)
Flow	RT	Area	RT	Area	RT	Area	RT	Area
	<sup>a</sup> 1,1 mL/min <sup>1</sup>				1,0 mL/min			
	257 nm		375 nm		257 nm		375 nm	
23 °C	0.42	0.40	0.44	0.40	0.55	0.50	0.58	0.57
25 °C	0.33	0.30	0.32	0.30	0.46	0.43	0.46	0.46
30 °C <sup>b</sup>	0.09	0.07	0.09	0.05	0.22	0.22	0.23	0.22
35 °C	0.08	0.08	0.10	0.11	0.15	0.15	0.15	0.15

<sup>a</sup> Working flow rate in the proposed method

<sup>b</sup> Working column temperature in the proposed method




Fig. 1. Chemical structure of quercetin.




Fig. 2. UV spectra of quercetin standard,  $c = 0.02$  mg/mL (a) and ascorbic acid,  $c = 0.02$  mg/mL (b).




Fig. 3. Typical chromatograms obtained at 257 nm and 375 nm from a mixed standard solution containing quercetin ( $c = 0.02$  mg/mL) and ascorbic acid ( $c = 0.02$  mg/mL).




Fig. 4. Typical chromatograms obtained at a 257 nm from sample solution (capsules containing 250 mg quercetin and 250 mg ascorbic acid) (a), quercetin standard solution (b), blank solution (c).




Fig. 5. Typical chromatograms obtained at a 375 nm from sample solution (capsules containing 250 mg quercetin and 250 mg ascorbic acid) (a), quercetin standard solution (b), blank solution (c).




Fig. 6. UV spectra of quercetin peak, eluted at retention time of 5.98 min from the standard solution (a) and sample solution of capsules containing 250 mg quercetin and 250 mg ascorbic acid (b).